


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 0407 Accredited to ISO/IEC 17025:2005	Campden Technology Ltd	
	Issue No: 022 Issue date: 04 February 2011	
	Chipping Campden Gloucestershire GL55 6LD	Contact: Mr M George Tel: +44 (0)1386 842000 Fax: +44 (0)1386 842100 E-Mail: info@campden.co.uk Website: www.campden.co.uk
Calibration performed at the above address only		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
TEMPERATURE Liquid-in-glass (including maximum thermometers), resistance thermometers and electronic thermometers with sensors	- 15 °C to 0 °C 0 °C to 130 °C	0.080 °C 0.080 °C	
END			