


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>UKAS CALIBRATION</p> <p>0407</p> <p>Accredited to ISO/IEC 17025:2005</p>	<p>CCFRA Technology Ltd</p> <p>Issue No: 017 Issue date: 09 March 2007</p>	
	<p>Chipping Campden Gloucestershire GL55 6LD</p>	<p>Contact: Mr R Thorn Tel: +44 (0)1386 842000 Fax: +44 (0)1386 842100 E-Mail: info@campden.co.uk Website: www.campden.co.uk</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks
TEMPERATURE			
Liquid-in-glass and resistance thermometers calibrated in saturated steam	105 °C to 130 °C	0.15 °C plus: See Note 1	NOTE 1 Analogue - 0.5 scale division Digital - one least significant digit
Liquid-in-glass and electronic thermometers with sensors	-15 °C to 0 °C 0 °C to 130 °C	0.07 °C plus: See Note 1 0.06 °C plus: See Note 1	
END			